



# CERTIFICATE OF ACCREDITATION

**The ANSI National Accreditation Board**

Hereby attests that

**Techmaster de Mexico SA de CV**  
Calle Seminario #8610 Int. 11,  
Col. Niños Heroes, deleg. La Presa. C.P. 22120  
Parque Industrial Arboledas, Tijuana, B.C., Mexico  
(and satellite locations as listed on the scope)

Fulfils the requirements of

**ISO/IEC 17025:2017**

and national standard

**ANSI/NCSL Z540-1-1994 (R2002)**

In the field of

**CALIBRATION, DIMENSIONAL MEASUREMENT and TESTING**

This certificate is valid only when accompanied by a current scope of accreditation document.

The current scope of accreditation can be verified at [www.anab.org](http://www.anab.org).

R. Douglas Leonard Jr., VP, PILR SBU

Expiry Date: 29 October 2024  
Certificate Number: AC-1342



This laboratory is accredited in accordance with the recognized International Standard ISO/IEC 17025:2017.  
This accreditation demonstrates technical competence for a defined scope and the operation of a laboratory quality management system (refer to joint ISO-ILAC-IAF Communiqué dated April 2017).



## SCOPE OF ACCREDITATION TO ISO/IEC 17025:2017

AND

**ANSI/NCSL Z540-1-1994 (R2002)**

**Techmaster de Mexico SA de CV**

Calle Seminario #8610 Int. 11

Col. Niños Heroes, deleg. La Presa. C.P. 22120, Parque Industrial Arboledas  
Tijuana, B.C., Mexico

J. Ernesto Matamoros      Phone: (52) 664-624-4444  
quality.mx@techmaster.us    www.techmasterdemexico.com

### Services performed at satellite locations as indicated

Bvd Lázaro Cárdenas 543 F03, Ex Ej. Coahuila, Mexicali, B.C. C.P. 21360.

Contact: Mauricio Garayzar. Tels: 01(686) 555-1660, 557-6117 [mexicali@techmaster.us](mailto:mexicali@techmaster.us)  
Ave Antonio J Bermúdez No. 2151-5, Fraccionamiento Parque Industrial Antonio J Bermúdez  
Código Postal 32470, Ciudad Juárez Chih

Phone 01(656) 687-2471,648-1181 [cdjuarez@techmaster.us](mailto:cdjuarez@techmaster.us)

Ave. Ignacio Morones Prieto No. 914 Ote. Int. 112 Col. La Huerta C.P. 67144 Guadalupe, N.L.  
Phone: 01(81)1334-0701 [monterrey@techmaster.us](mailto:monterrey@techmaster.us)

Carretera Federal 57, Mexico-Queretaro Lateral Norte Km 201+100 Interior 26, El Marques  
(Business Park Conin) Queretaro, Queretaro  
Phone :01(442)241-5104 [queretaro@techmaster.us](mailto:queretaro@techmaster.us)

## CALIBRATION, DIMENSIONAL MEASUREMENT AND TESTING

Valid to: October 29, 2024

Certificate Number: AC-1342

### CALIBRATION

#### Acoustics and Vibration

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Sound - Source	(94, 114) dB (251, 1 000) Hz	0.29 dB	Sound Calibrator Tijuana Mexicali Juarez Monterrey Queretaro

## Acoustics and Vibration

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Sound - Measure	(30 to 140) dB	0.43 dB	Sound Level Meter Tijuana Mexicali Juarez Monterrey Queretaro
Vibration - Source	(0.2 to 20) gpk (7 to 100) Hz 100 Hz to 5 kHz (5 to 10) kHz	3.5 % of reading + 0.03 g 3.1 % of reading + 0.03 g 3.6 % of reading + 0.03 g	Portable Vibration Calibrator Tijuana Mexicali Juarez Monterrey Queretaro
Vibration - Measure	Up to 500 g (1 to 10 Hz) (10 Hz to 2kHz) (2 kHz to 10 kHz)	3.5 % of reading + 0.06 g 2.3 % of reading + 0.06 g 4.4 % of reading + 0.06 g	Portable Vibration Meter Tijuana Mexicali Juárez Monterrey Queretaro

## Chemical Quantities

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
pH measuring equipment	1.68 pH 4.0 pH 7.0 pH 10 pH 11 pH	0.02 pH 0.02 pH 0.02 pH 0.04 pH 0.02 pH	pH Solutions Tijuana Mexicali Juarez Monterrey Queretaro
Viscosity Dynamic measuring equipment <sup>4</sup>	10 mPa·s (cP) 100 mPa·s (cP) 1 000 mPa·s (cP) 5 000 mPa·s (cP) 12 500 mPa·s (cP) 100 000 mPa·s (cP) 200 000 mPa·s (cP)	0.26 mPa·s (cP) 1.1 mPa·s (cP) 5.1 mPa·s (cP) 7.4 mPa·s (cP) 48 mPa·s (cP) 80 mPa·s (cP) 210 mPa·s (cP)	Standard Solutions: S6, S60, D500, N350, S2000, D7500, S8000 Tijuana Mexicali Juarez Monterrey Queretaro

### Chemical Quantities

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Conductivity measuring equipment <sup>4</sup>	2 $\mu\text{mhos}/\text{cm}$ 10 $\mu\text{mhos}/\text{cm}$ 100 $\mu\text{mhos}/\text{cm}$ 1 000 $\mu\text{mhos}/\text{cm}$ 1 400 $\mu\text{mhos}/\text{cm}$ 10 000 $\mu\text{mhos}/\text{cm}$ 100 000 $\mu\text{mhos}/\text{cm}$	0.25 $\mu\text{mhos}/\text{cm}$ 1.2 $\mu\text{mhos}/\text{cm}$ 15 $\mu\text{mhos}/\text{cm}$ 120 $\mu\text{mhos}/\text{cm}$ 180 $\mu\text{mhos}/\text{cm}$ 2 200 $\mu\text{mhos}/\text{cm}$ 8 000 $\mu\text{mhos}/\text{cm}$	Conductivity Solutions Tijuana Mexicali Juarez Monterrey Queretaro

### Electrical – DC/Low Frequency

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
DC Voltage - Source	(2.2 to 220) mV 200 mV to 2.2 V (2.2 to 11) V (11 to 22) V (22 to 220) V 220 V to 1.1 kV	11 $\mu\text{V}/\text{V} + 0.48 \mu\text{V}$ 6.2 $\mu\text{V}/\text{V} + 0.87 \mu\text{V}$ 4.2 $\mu\text{V}/\text{V} + 3 \mu\text{V}$ 4.2 $\mu\text{V}/\text{V} + 5.2 \mu\text{V}$ 6.1 $\mu\text{V}/\text{V} + 99 \mu\text{V}$ 8 $\mu\text{V}/\text{V} + 0.53 \text{ mV}$	Multiproduct Calibrator Tijuana Mexicali Juarez Monterrey Queretaro
DC Voltage - Measure	(2 to 200) mV 200 mV to 2 V (2 to 20) V (20 to 200) V 200 V to 1 kV	6.7 $\mu\text{V}/\text{V} + 0.2 \mu\text{V}$ 4.3 $\mu\text{V}/\text{V} + 0.5 \mu\text{V}$ 4.3 $\mu\text{V}/\text{V} + 4.8 \mu\text{V}$ 6.7 $\mu\text{V}/\text{V} + 98 \mu\text{V}$ 6.7 $\mu\text{V} + 0.63 \text{ mV}$	Reference Multimeter Tijuana Mexicali Juarez Monterrey Queretaro
DC High Voltage - Measure	Up to 10 kV (10 to 100) kV	0.35 mV/V + 0.09 V 0.63 mV/V + 4.1 V	High Voltage Meter Tijuana Mexicali Juarez Monterrey Queretaro
Charge Analyzer	Up to 1 kV (1 to 5) kV	24 mV/V + 0.5 V 24 mV/V + 12 V	Charge Plate Analyzer Tijuana Mexicali Juarez Monterrey Queretaro

## Electrical – DC/Low Frequency

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
DC Current - Source	(2 to 220) $\mu$ A 220 $\mu$ A to 2.2 mA (2.2 to 22) mA (22 to 220) mA 220 mA to 2.2 A	50 $\mu$ A/A + 7.2 nA 43 $\mu$ A/A + 8.4 nA 43 $\mu$ A/A + 48 nA 55 $\mu$ A/A + 0.84 $\mu$ A 97 $\mu$ A/A + 21 $\mu$ A	Multiproduct Calibrator Tijuana Mexicali Juarez Monterrey Queretaro
DC Current - Source	(2.2 to 10) A (10 to 10.9) A (10.9 to 20.5) A	0.6 mA/A + 0.77 mA 0.6 mA/A + 1 mA 1.2 mA/A + 1.4 mA	Multiproduct Calibrator Tijuana Mexicali Juarez Monterrey Queretaro
DC Current – Source Clamp On Meters	(10 to 550) A (550 to 1 025) A	2.5 mA/A + 0.55 A 2.6 mA/A + 0.55 A	Multiproduct Calibrator with 50 Turn Coil Tijuana Mexicali Juarez Monterrey Queretaro
DC Current - Measure	(2 to 200) $\mu$ A (200 $\mu$ A to 2) mA (2 to 20) mA (20 to 200) mA 200 mA to 2 A (2 to 20) A	15 $\mu$ A/A + 0.49 nA 15 $\mu$ A/A + 4.8 nA 17 $\mu$ A/A + 48 nA 59 $\mu$ A/A + 0.96 $\mu$ A 0.23 mA/A + 25 $\mu$ A 0.49 mA/A + 0.36 mA	Reference Multimeter Tijuana Mexicali Juarez Monterrey Queretaro
DC Current – Measure	(20 to 100) A (100 to 300) A	0.5 mA/A + 0.004 A 1 mA/A + 0.004 A	Current Shunt Tijuana Mexicali Juarez Monterrey Queretaro
DC Current - Measure	Up to 1 000 A	2.6 mA/A + 20 mA	Current Shunt Tijuana Mexicali Juarez Monterrey Queretaro

## Electrical – DC/Low Frequency

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Resistance - Source (Fixed Values)	0.001 Ω 0.01 Ω 0.1 Ω 0.333 Ω	0.23 mΩ 0.22 mΩ 0.2 mΩ 0.9 mΩ	Reference Resistor Tijuana Mexicali Juarez Monterrey Queretaro
Resistance - Source (Fixed Values)	1 Ω 1.9 Ω 10 Ω 19 Ω 100 Ω 190 Ω	0.12 mΩ 0.22 mΩ 0.28 mΩ 0.53 mΩ 1.3 mΩ 2.3 mΩ	Multiproduct Calibrator Tijuana Mexicali Juarez Monterrey Queretaro
Resistance - Source (Fixed Values)	1 kΩ 1.9 kΩ 10 kΩ 19 kΩ 100 kΩ 190 kΩ 1 MΩ 1.9 MΩ 10 MΩ 19 MΩ 100 MΩ	11 mΩ 20 mΩ 0.11 Ω 0.2 Ω 1.4 Ω 2.6 Ω 25 Ω 51 Ω 0.5 kΩ 1.2 kΩ 14 kΩ	Multiproduct Calibrator Tijuana Mexicali Juarez Monterrey Queretaro
Resistance - Source	100 MΩ to 1 GΩ (1 to 10) GΩ 10 GΩ to 1 TΩ	5.1% of reading + 5 MΩ 20 % of reading + 20 MΩ 20 % of reading + 22 MΩ	Decade Resistor Tijuana Mexicali Juarez Monterrey Queretaro
Resistance - Measure	Up to 2Ω (2 to 20) Ω (20 to 200) Ω 200 Ω to 2 kΩ (2 to 20) kΩ (20 to 200) kΩ 200 kΩ to 2 MΩ (2 to 20) MΩ (20 to 200) MΩ 200 MΩ to 2 GΩ	23 μΩ/Ω + 5.9 μΩ 12 μΩ/Ω + 18 μΩ 9.8 μΩ/Ω + 80 μΩ 12 μΩ/Ω + 0.94 mΩ 11 μΩ/Ω + 47 mΩ 11 μΩ/Ω + 60 mΩ 13 μΩ/Ω + 1.2 Ω 28 μΩ/Ω + 0.12 kΩ 0.15 mΩ/Ω + 1.2 kΩ 1.9 mΩ/Ω + 12 kΩ	Reference Multimeter Tijuana Mexicali Juarez Monterrey Queretaro

## Electrical – DC/Low Frequency

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Resistance – Measure at 500 V	200 kΩ to 10 GΩ	61 mΩ/Ω + 0.6 MΩ	
at 500 V	(10 to 100) GΩ	0.24 Ω/Ω + 0.6 MΩ	
at 1 kV	200 kΩ to 20 GΩ	60 mΩ/Ω + 0.6 MΩ	
at 1 kV	(20 to 200) GΩ	0.24 Ω/Ω + 0.6 MΩ	
at 2.5 kV	200 kΩ to 50 GΩ	60 mΩ/Ω + 0.6 MΩ	
at 5 kV	200 kΩ to 100 GΩ	60 mΩ/Ω + 0.6 MΩ	
at 5 kV	100 GΩ to 1 TΩ	0.24 Ω/Ω + 0.6 MΩ	
at 10 kV	200 kΩ to 200 GΩ	62 mΩ/Ω + 0.6 MΩ	
at 10 kV	200 GΩ to 2 TΩ	0.24 Ω/Ω + 0.6 MΩ	
AC Voltage - Source	Up to 2.2 mV	0.29 mV/V + 4.8 μV	
	(10 to 20) Hz	0.11 mV/V + 4.8 μV	
	(20 to 40) Hz	96 μV/V + 9.6 μV	
	40 Hz to 20 kHz	0.24 mV/V + 4.8 μV	
	(20 to 50) kHz	0.6 mV/V + 6 μV	
	(50 to 100) kHz	1.3 mV/V + 12 μV	
	(100 to 300) kHz	1.7 mV/V + 24 μV	
	(300 to 500) kHz	3.3 mV/V + 24 μV	
	500 kHz to 1 MHz	0.3 mV/V + 4.8 μV	Multiproduct Calibrator
	(2.2 to 22) mV	0.13 mV/V + 4.8 μV	
	(10 to 20) Hz	0.12 mV/V + 4.8 μV	
	(20 to 40) Hz	0.24 mV/V + 4.8 μV	
	40 Hz to 20 kHz	0.62 mV/V + 6 μV	
	(20 to 50) kHz	1.3 mV/V + 12 μV	
	(50 to 100) kHz	1.7 mV/V + 24 μV	
	(100 to 300) kHz	3.3 mV/V + 24 μV	
	(300 to 500) kHz	0.29 mV/V + 20 μV	
	500 kHz to 1 MHz	0.11 mV/V + 8.9 μV	
	(22 to 220) mV	98 μV/V + 8.5 μV	
	(10 to 20) Hz	0.24 mV/V + 8.5 μV	
	(20 to 40) Hz	0.55 mV/V + 21 μV	
	40 Hz to 20 kHz	1.1 mV/V + 24 μV	
	(20 to 50) kHz	1.7 mV/V + 32 μV	
	(50 to 100) kHz	3.3 mV/V + 54 μV	

## Electrical – DC/Low Frequency

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
AC Voltage - Source	220 mV to 2.2 V (10 to 20) Hz (20 to 40) Hz 40Hz to 20 kHz (20 to 50) kHz (50 to 100) kHz (100 to 300) kHz (300 to 500) kHz 500 kHz to 1 MHz (2.2 to 22) V (10 to 20) Hz (20 to 40) Hz 40 Hz to 20 kHz (20 to 50) kHz (50 to 100) kHz (100 to 300) kHz (300 to 500) kHz 500 kHz to 1 MHz (22 to 220) V (10 to 20) Hz (20 to 40) Hz 40 Hz to 20 kHz (20 to 50) kHz (50 to 100) kHz (100 to 300) kHz (300 to 500) kHz (500 kHz to 1 MHz) 220 V to 1.1 kV (15 to 50) Hz 50 Hz to 1 kHz	0.29 mV/V + 48 µV 0.11 mV/V + 18 µV 57 µV/V + 9.6 µV 91 µV/V + 12 µV 0.13 mV/V + 36 µV 0.5 mV/V + 96 µV 1.2 mV/V + 0.24 mV 2.1 mV/V + 0.36 mV  0.29 mV/V + 0.14 mV 0.11 mV/V + 41 µV 57 µV/V + 23 µV 91 µV/V + 21 µV 0.13 mV/V + 43 µV 1.2 mV/V + 0.25 mV 2.1 mV/V + 0.4 mV  0.29 mV/V + 0.12 V 0.12 mV/V + 1.9 mV 75 µV/V + 0.72 mV 0.10 mV/V + 1.2 mV 0.19 mV/V + 3 mV 1.1 mV/V + 19 mV 5.3 mV/V + 48 mV 9.7 mV/V + 96 mV  0.29 mV/V + 48 mV 0.11 mV/V + 18 mV	Multiproduct Calibrator  Tijuana Mexicali Juarez Monterrey Queretaro
AC Voltage - Measure	Up to 200 mV (1 to 10) Hz (10 to 40) Hz (40 to 100) Hz 100 Hz to 2 kHz (2 to 10) kHz (10 to 30) kHz (30 to 100) kHz	0.2 mV/V + 29 µV 0.17 mV/V + 5 µV 0.14 mV/V + 5 µV 0.17 mV/V + 2 µV 0.17 mV/V + 5 µV 0.42 mV/V + 10 µV 0.93 mV/V + 24 µV	Reference Multimeter  Tijuana Mexicali Juarez Monterrey Queretaro

## Electrical – DC/Low Frequency

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
AC Voltage - Measure	200 mV to 2 V (1 to 10) Hz (10 to 40) Hz (40 to 100) Hz 100 Hz to 2 kHz (2 to 10) kHz (10 to 30) kHz (30 to 100) kHz (2 to 20) V (1 to 10) Hz (10 to 40) Hz (40 to 100) Hz 100 Hz to 2 kHz (2 to 10) kHz (10 to 30) kHz (30 to 100) kHz (100 to 300) kHz 300 kHz to 1 MHz (20 to 200) V 1 to 10 Hz (10 to 40) Hz (40 to 100) Hz 100 Hz to 2 kHz (2 to 10) kHz (10 to 30) kHz (30 to 100) kHz (100 to 300) kHz 300 kHz to 1 MHz 200 V to 1 kV (1 to 10) Hz (10 to 40) Hz 40 Hz to 10 kHz (10 to 30) kHz (30 to 100) kHz	0.19 mV/V + 0.33 mV 0.15 mV/V + 3 µV 0.12 mV/V + 24 µV 0.17 mV/V + 24 µV 0.14 mV/V + 24 µV 0.27 mV/V + 48 µV 0.7 mV/V + 0.24 mV  0.18 mV/V + 3.3 mV 0.14 mV/V + 0.3 mV 0.11 mV/V + 0.24 mV 0.15 mV/V + 0.24 mV 0.14 mV/V + 0.24 mV 0.27 mV/V + 0.48 mV 0.69 mV/V + 2.4 mV 3.6 mV/V + 24 mV 12 mV/V + 0.24 mV  0.19 mV/V + 58 mV 0.15 mV/V + 2.7 mV 0.12 mV/V + 2.4 mV 0.15 mV/V + 2.4 mV 0.14 mV/V + 2.4 mV 0.27 mV/V + 4.8 mV 0.69 mV/V + 24 mV 3.6 mV/V + 0.24 mV 12 mV/V + 2.4 V  0.19 mV/V + 84 mV 0.16 mV/V + 24 mV 0.29 mV/V + 24 mV 0.28 mV/V + 48 mV 0.72 mV/V + 0.24 V	Reference Multimeter  Tijuana Mexicali Juarez Monterrey Queretaro
AC Voltage - Measure	Up to 10 kV (30 to 200) Hz (200 to 450) Hz (450 to 600) Hz (10 to 100) kV (30 + 70) Hz (70 to 200) Hz	1.4 mV/V + 0.14 V 4.6 mV/V + 0.14V 8.7 mV/V + 0.14V  1.4 mV/V + 0.7 V 17 % + 0.7 V	High Voltage Meter  Tijuana Mexicali Juarez Monterrey Queretaro

## Electrical – DC/Low Frequency

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
AC Current - Source	Up to 220 $\mu$ A (10 to 20) Hz (20 to 40) Hz 40 Hz to 1 kHz (1 to 5) kHz (5 to 10) kHz  220 $\mu$ A to 2.2 mA (10 to 20) Hz (20 to 40) Hz 40 Hz to 1 kHz (1 to 5) kHz (5 to 10) kHz  (2.2 to 22) mA (10 to 20) Hz (20 to 40) Hz 40 Hz to 1 kHz (1 to 5) kHz (5 to 10) kHz  (22 to 220) mA (10 to 20) Hz (20 to 40) Hz 40 Hz to 1 kHz (1 to 5) kHz (5 to 10) kHz  220 mA to 2.2 A 20 Hz to 1 kHz (1 to 5) kHz (5 to 10) kHz	0.38 mA/A + 21 nA 0.3 mA/A + 12 nA 0.27 mA/A + 9.9 nA 0.41 mA/A + 15 nA 1.4 mA/A + 78 nA  0.3 mA/A + 68 nA 0.19 mA/A + 48 nA 0.15 mA/A + 48 nA 0.27 mA/A + 0.13 $\mu$ A 1.3 mA/A + 0.78 $\mu$ A  0.3 mA/A + 0.49 $\mu$ A 0.2 mA/A + 0.43 $\mu$ A 0.15 mA/A + 0.43 $\mu$ A 0.39 mA/A + 0.66 $\mu$ A 1.4 mA/A + 6 $\mu$ A  0.43 mA/A + 7.9 $\mu$ A 0.37 mA/A + 4.7 $\mu$ A 0.36 mA/A + 3.9 $\mu$ A 0.41 mA/A + 4.3 $\mu$ A 3.3 mA/A + 12 $\mu$ A  0.5 mA/A + 55 $\mu$ A 4 $\mu$ A/A + 96 $\mu$ A 9.3 mA/A + 0.19 mA	Multiproduct Calibrator  Tijuana Mexicali Juarez Monterrey Queretaro
AC Current – Source	(3 to 20) A (45 to 100) Hz 100 Hz to 1 kHz (1 to 5) kHz  (11 to 20.5) A (45 to 100) Hz 100 Hz to 1 kHz (1 to 5) kHz	0.5 mA/A + 1.4 mA 0.7 mA/A + 1.4 mA 20 mA/A + 1.4 mA  0.8 mA/A + 3.4 mA 1 mA/A + 3.4 mA 20 mA/A + 3.4 mA	Multiproduct Calibrator  Tijuana Mexicali Juarez Monterrey Queretaro

## Electrical – DC/Low Frequency

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
AC Current – Source Clamp On Meters	(16.5 to 55) A 65 Hz 440 Hz (55 to 150) A 65 Hz 440 Hz (150 to 550) A 65 Hz 440 Hz	2.8 mA/A + 0.17 A 7.9 mA/A + 0.19 A  2.8 mA/A + 0.31 A 7.9 mA/A + 0.16 A  2.8 mA/A + 1.3 A 7.9 mA/A + 0.41 A	Multiproduct Calibrator with 50 Turn Coil Tijuana Mexicali Juarez Monterrey Queretaro
AC Current – Source Clamp On Meters	(550 to 1 025) A 65 Hz 440 Hz	2.9 mA/A + 0.71 A 8 mA/A + 1.2 A	Multiproduct Calibrator with 50 Turn Coil Tijuana Mexicali Juarez Monterrey Queretaro
AC Current - Measure	Up to 200 $\mu$ A 10 Hz to 10 kHz (10 to 30) kHz (30 to 100) kHz 200 $\mu$ A to 2 mA 10 Hz to 10 kHz (10 to 30) kHz (30 to 100) kHz 2 to 20 mA 10 Hz to 10 kHz (10 to 30) kHz (30 to 100) kHz 20 to 200 mA 10 Hz to 10 kHz (10 to 30) kHz 200 mA to 2 A 10 Hz to 2 kHz (2 to 10) kHz (10 to 30) kHz 2 to 20 A 10 Hz to 2 kHz (2 to 10) kHz	0.42 mA/A + 24 nA 0.88 mA/A + 24 nA 4.8 mA/A + 24 nA  0.36 mA/A + 0.24 $\mu$ A 0.86 mA/A + 0.24 $\mu$ A 4.8 mA/A + 0.24 $\mu$ A  0.37 mA/A + 2.4 $\mu$ A 0.86 mA/A + 2.4 $\mu$ A 4.8 mA/A + 2.4 $\mu$ A  0.35 mA/A + 24 $\mu$ A 0.75 mA/A + 24 $\mu$ A  0.75 mA/A + 0.25 mA 0.87 mA/A + 0.27 mA 3.6 mA/A + 0.26 mA  0.99 mA/A + 2.4 mA 3.1 mA/A + 2.4 mA	Reference Multimeter Tijuana Mexicali Juarez Monterrey Queretaro

## Electrical – DC/Low Frequency

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
AC Current - Measure	(20 to 1 000) A (60 to 100) Hz	2.6 mA/A + 0.28 A	Current Shunt Tijuana Mexicali Juarez Monterrey Queretaro
Electrical Simulation of Thermocouple Indicators	Type B  (600 to 800) °C (800 to 1 000) °C (1 000 to 1 550) °C (1 550 to 1 820) °C  Type C  (0 to 150) °C (150 to 650) °C (650 to 1 000) °C (1 000 to 1 800) °C (1 800 to 2 316) °C  Type E  (-250 to -100) °C (-100 to -25) °C (-25 to 350) °C (350 to 650) °C (650 to 1 000) °C  Type J  (-200 to -100) °C (-100 to -30) °C (-30 to 150) °C (150 to 760) °C (760 to 1 200) °C  Type K  (-200 to -100) °C (-100 to -25) °C (-25 to 120) °C (120 to 1 000) °C (1 000 to 1 372) °C  Type L  (-200 to -100) °C (-100 to 800) °C (800 to 900) °C	0.53 °C 0.43 °C 0.37 °C 0.41 °C  0.36 °C 0.30 °C 0.37 °C 0.61 °C 1 °C  0.6 °C 0.19 °C 0.17 °C 0.19 °C 0.26 °C  0.32 °C 0.19 °C 0.17 °C 0.21 °C 0.28 °C  0.57 °C 0.22 °C 0.19 °C 0.31 °C 0.48 °C  0.45 °C 0.32 °C 0.21 °C	Multiproduct Calibrator Tijuana Mexicali Juarez Monterrey Queretaro

## Electrical – DC/Low Frequency

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Electrical Simulation of Thermocouple Indicators	Type N (-200 to -100) °C (-100 to -25) °C (-25 to 120) °C (120 to 410) °C (410 to 1 300) °C  Type R (0 to 250) °C (250 to 400) °C (400 to 1 000) °C (1 000 to 1 767) °C  Type S (0 to 250) °C (250 to 400) °C (400 to 1 000) °C (1 000 to 1 767) °C  Type T (-250 to -150) °C (-150 to 0) °C (0 to 120) °C (120 to 400) °C  Type U (0 to 250) °C (250 to 1 000) °C (1 000 to 1 400) °C (1 400 to 1 767) °C	0.48 °C 0.26 °C 0.23 °C 0.22 °C 0.33 °C  0.69 °C 0.42 °C 0.48 °C 0.6 °C  0.69 °C 0.42 °C 0.48 °C 0.6 °C  0.76 °C 0.29 °C 0.19 °C 0.48 °C  0.57 °C 0.43 °C 0.44 °C 0.57 °C	Multiproduct Calibrator  Tijuana Mexicali Juarez Monterrey Queretaro
Electrical Simulation of RTD Indicators	Pt 385, 100 Ω (-200 to 0) °C (0 to 100) °C (100 to 300) °C (300 to 400) °C (400 to 630) °C (630 to 800) °C  Pt 385, 200 Ω (-200 to 100) °C (100 to 260) °C (260 to 300) °C (300 to 400) °C (400 to 600) °C (600 to 630) °C	0.06 °C 0.08 °C 0.11 °C 0.12 °C 0.14 °C 0.28 °C  0.05 °C 0.06 °C 0.14 °C 0.16 °C 0.17 °C 0.19 °C	Multiproduct Calibrator  Tijuana Mexicali Juarez Monterrey Queretaro

## Electrical – DC/Low Frequency

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment	
Electrical Simulation of RTD Indicators	Pt 385, 500 Ω (-200 to -80) °C (-80 to 100) °C (100 to 260) °C (260 to 400) °C (400 to 600) °C (600 to 630) °C  Pt 385, 1 000Ω (-200 to 0) °C (0 to 100) °C (100 to 260) °C (260 to 300) °C (300 to 600) °C (600 to 630) °C  Pt 3926, 100 Ω (-200 to 0) °C (0 to 100) °C (100 to 300) °C (300 to 400) °C (400 to 630) °C	0.07 °C 0.07 °C 0.08 °C 0.11 °C 0.12 °C 0.14 °C  0.06 °C 0.07 °C 0.07 °C 0.08 °C 0.1 °C 0.28 °C  0.06 °C 0.08 °C 0.11 °C 0.12 °C 0.14 °C	Multiproduct Calibrator  Tijuana Mexicali Juarez Monterrey Queretaro	
Capacitance - Source	10 Hz to 10 kHz 10 Hz to 10 kHz 10 Hz to 3 kHz 10 Hz to 1 kHz 10 Hz to 1 kHz 10 Hz to 1 kHz 10 Hz to 1 kHz (10 to 600) Hz (10 to 300) Hz (10 to 150) Hz (10 to 120) Hz (10 to 80) Hz	(200 to 400) pF 400 pF to 1.1 nF (1.1 to 3.3) nF (3.3 to 11) nF (11 to 33) nF (33 to 110) nF (110 to 330) nF 330 nF to 1.1 μF (1.1 to 3.3) μF (3.3 to 11) μF (11 to 33) μF (33 to 110) μF	6 mF/F + 12 pF 6 mF/F + 12 pF 6 mF/F + 12 pF 3 mF/F + 13 pF 3 mF/F + 0.12 nF 3 mF/F + 0.14 nF 3 mF/F + 0.43 nF 3 mF/F + 1.5 nF 3 mF/F + 4.1 nF 3 mF/F + 14 nF 4.8 mF/F + 41 nF 5.4 mF/F + 0.16 μF	Multiproduct Calibrator  Tijuana Mexicali Juarez Monterrey Queretaro
Capacitance - Measure	(1 to 100) kHz 300 Hz to 100 kHz (50 to 100) kHz (50 to 200) kHz 50 Hz to 10 kHz 50 Hz to 1 kHz	100 pF to 1 nF (1 to 10) nF (10 to 100) nF (0.1 to 1) μF (1 to 10) μF (10 to 100) μF	1.3 mF/F + 0.013 pF 1.3 mF/F + 0.17 pF 1.2 mF/F + 2.7 pF 1.2 mF/F + 0.14 nF 1.2 mF/F + 1.1 nF 1.2 mF/F + 1.7 nF	Impedance Meter Tijuana Mexicali Juarez Monterrey Queretaro

## Electrical – DC/Low Frequency

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Capacitance - Source (Fixed Values @ 1 kHz)	1 nF 10 nF 100 nF 1 $\mu$ F	0.8 pF 14 pF 0.14 nF 0.65 nF	Reference Capacitors Tijuana Mexicali Juarez Monterrey Queretaro
Inductance - Source 100 Hz to 10 kHz	100 $\mu$ H 1 mH 20 mH 100 mH @ (0.1 to 1) kHz 101.88 mH @ 10 kHz	0.85 $\mu$ H 18 $\mu$ H 27 $\mu$ H 0.13 mH 0.15 mH	Reference Inductor Tijuana Mexicali Juarez Monterrey Queretaro
Inductance - Measure (2 to 100) kHz 300 Hz to 100k Hz 100 Hz to 100 kHz (50 to 100) kHz 50 Hz to 10 kHz 50 Hz to 2 kHz	100 $\mu$ H to 1 mH (1 to 10) mH (10 to 100) mH 100 mH to 1 H (1 to 10) H (10 to 100) H	1.2 mH/H + 23 nH 1.2 mH/H + 0.27 $\mu$ H 1.2 mH/H + 3.3 $\mu$ H 1.3 mH/H + 27 $\mu$ H 1.3 mH/H + 1.1 $\mu$ H 1.3 mH/H + 3.5 mH	Impedance Meter Tijuana Mexicali Juarez Monterrey Queretaro
Oscilloscopes Amplitude Square Wave 50 $\Omega$ Load	1 mV to 6.6 V p-p 10 Hz to 10 kHz	3 mV/V + 0.96 mV	
1 M $\Omega$ Load	1 mV to 130 V p-p 10 Hz to 10 kHz	3 mV/V + 1.8 mV	Oscilloscope Calibrator Tijuana Mexicali Juarez Monterrey Queretaro
Leveled Sine Wave	5 mV to 5.5 V 50 kHz to 100 MHz (100 to 300) MHz (300 to 500) MHz (500 to 600) MHz	42 mV/V + 1.4 mV 48 mV/V + 1.4 mV 66 mV/V + 1.4 mV 72 mV/V + 1.4 mV	
Time Marker into 50 $\Omega$	1 ns to 50 ms 50 ms to 5 s	1 $\mu$ s/s + 60 ns 3 $\mu$ s/s + 9 $\mu$ s	
DC Power - Source	Up to 3.06 kW (3.06 to 20.91) kW	20 $\mu$ W/W + 0.39 W 50 $\mu$ W/W + 3.1W	Multiproduct Calibrator Tijuana Mexicali Juarez Monterrey Queretaro

**Electrical – DC/Low Frequency**

<b>Parameter/Equipment</b>	<b>Range</b>	<b>Expanded Uncertainty of Measurement (+/-)</b>	<b>Reference Standard, Method, and/or Equipment</b>
AC Power - Source (45 to 65 Hz)	Up to 336.6 W 336.6 W to 2.244 kW (2.244 to 4.59) kW (4.59 to 20.91) kW	50 $\mu$ W/W + 0.39 W 60 $\mu$ W/W + 2.8 W 90 $\mu$ W/W + 2.8 W 50 $\mu$ W/W + 2.8 W	Multiproduct Calibrator Tijuana Mexicali Juarez Monterrey Queretaro
Phase	Up to 90 ° (10 to 500) Hz 500 Hz to 1 kHz (1 to 5) kHz (5 to 10) kHz (10 to 30) kHz	1.4 ° 1.5 ° 2.2 ° 3.7 ° 6.9 °	Multiproduct Calibrator Tijuana Mexicali Juarez Monterrey Queretaro

**Electrical – RF/Microwave**

<b>Parameter/Equipment</b>	<b>Range</b>	<b>Expanded Uncertainty of Measurement (+/-)</b>	<b>Reference Standard, Method, and/or Equipment</b>
RF Power/Gain – Measure <sup>1</sup>	(-30 to 20) dB (10 to 20) MHz (20 to 50) MHz (50 to 100) MHz 100 MHz to 1 GHz (1 to 4) GHz (4 to 8) GHz (8 to 18) GHz	2.1 % of reading + 0.09 dB 1.8 % of reading + 0.09 dB 1.4 % of reading + 0.09 dB 1.2 % of reading + 0.09 dB 1.2 % of reading + 0.09 dB 1.4 % of reading + 0.09 dB 2.5 % of reading + 0.09 dB	Feed thru Power Standard, Control Unit Tijuana Mexicali Juarez Monterrey Queretaro
Frequency Modulation - Measure	Rate: 20 Hz to 10 kHz Deviation: ≤ 40 kHz peak 250 kHz to 10 MHz  Rate: 20 Hz to 10 kHz Deviation: ≤ 400 kHz peak 10 MHz to 1.3 GHz	2.4 % of reading + 210 Hz  1.2 % of reading + 210 Hz	Measuring Receiver w/ Power Sensor Tijuana Mexicali Juarez Monterrey Queretaro
Amplitude Modulation - Measure	Rate: 50 Hz to 10 kHz Depths; (5 to 99) % 150 kHz to 10 MHz  Rate 10 MHz to 1.3 GHz Depths (5 to 99) % 50 Hz to 50 kHz	2.4 % of reading + 0.19 % depth  1.2 % of reading + 0.19 % depth	Measuring Receiver w/ Power Sensor Tijuana Mexicali Juarez Monterrey Queretaro

## Electrical – RF/Microwave

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Phase Modulation - Measure	Rate: 200 Hz to 10 kHz 150 kHz to 10 MHz  Rate: 200 Hz to 20 kHz 10 MHz to 1.3 GHz	4.8 % of reading + 0.32 rad  3.6 % of reading + 0.32 rad	Measuring Receiver w/ Power Sensor Tijuana Mexicali Juarez Monterrey Queretaro
RF Power - Measure	(-20 to 30) dBm 100 kHz to 2.6 GHz	0.1 dB	Measuring Receiver w/ Power Sensor Tijuana Mexicali Juarez Monterrey Queretaro
RF Power - Measure	(-30 to 20) dBm 100 kHz to 4.2 GHz 50 MHz to 26.5 GHz	4.9 % of reading + 0.21 dB 3.1 % of reading + 0.13 dB	Power Sensors w/Power Meter Tijuana Mexicali Juarez Monterrey Queretaro
Tuned RF Power Relative - Measure	2.5 MHz to 1.3 GHz (0 to -10) dB (-10 to -40) dB (-40 to -50) dB (-50 to -80) dB (-80 to -90) dB (-90 to -110) dB (-110 to -127) dB	0.03 dB 0.06 dB 0.13 dB 0.18 dB 0.16 dB 0.43 dB 0.44 dB	Measuring Receiver w/ Power Sensor Tijuana Mexicali Juarez Monterrey Queretaro
RF Power - Source	10 MHz to 2 GHz (13 to 10) dBm (10 to -10) dBm (-10 to -60) dBm (-60 to -110) dBm  2 to 20 GHz (13 to 10) dBm (10 to -10) dBm (-10 to -60) dBm (-60 to -110) dBm	1.5 dB 0.73 dB 1.1 dB 1.7 dB  1.6 dB 0.84 dB 1.2 dB 1.8 dB	Signal Generator Tijuana Mexicali Juarez Monterrey Queretaro

## Electrical – RF/Microwave

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
RF Power - Source	20 GHz to 26.5 GHz (13 to- 10) dBm (-10 to -60) dBm (-60 to -1) dBm	1.1 dB 1.5 dB 1.8 dB	Signal Generator Tijuana Mexicali Juarez Monterrey Queretaro

## Length – Dimensional Metrology

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Plug Gages <sup>1,2</sup>	Up to 0.5 in (0.5 to 1) in (1 to 2.5) in (2.5 to 4) in (4 to 12) in	(6.5 + 2.6L) $\mu$ in (7.8 + 2.3L) $\mu$ in (11 + 4.6L) $\mu$ in (18 + 3L) $\mu$ in (13 + 7.7L) $\mu$ in	LabMaster System and Gage Blocks Tijuana Juarez Monterrey
Ring Gages <sup>1,2</sup>	Up to 1 in (1 to 4) in (4 to 10) in (10 to 40) in	(9 + 20L) $\mu$ in (8.1 + 8.4L) $\mu$ in (11 + 11L) $\mu$ in [26 + 13 (L-10)] $\mu$ in	LabMaster System and Gage Blocks Tijuana Juarez Monterrey
Gage Blocks <sup>1,2</sup>	Up to 1 in (1 to 4) in (4 to 10) in	(4 + 0.8L) $\mu$ in (3.1 + 1.7L) $\mu$ in (1.2 + 2.2L) $\mu$ in	LabMaster System and Gage Blocks Tijuana Juarez
Gage Blocks <sup>1,2</sup>	Up to 1 in (1 to 4) in (4 to 10) in (10 to 40) in	(4 + 0.8L) $\mu$ in (3.1 + 1.7L) $\mu$ in (1.2 + 2.2L) $\mu$ in [9.7 + 13 (L-10)] $\mu$ in	LabMaster System and Gage Blocks Monterrey
Gage Blocks <sup>1,2</sup>	Up to 4 in Up to 101.6 mm	(3.9 + 0.42 L) $\mu$ in (0.12 + 0.011 L) $\mu$ m	Electromechanical Comparator & Gage Blocks Mexicali
Thread Plug Gages <sup>1,2</sup> Major Diameter and Pitch Diameter	(0.05 to 1) in (1 to 2) in (2 to 12) in	100 $\mu$ in 100 $\mu$ in 130 $\mu$ in	LabMaster System Gage Blocks and Thread Wires Tijuana Juarez Monterrey

## Length – Dimensional Metrology

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Thread Plug Gages <sup>1,2</sup> Major Diameter Pitch Diameter Flank Angle	Up to 5.9 in Up to 5.9 in (27 to 80) <sup>o</sup>	(76 + 5.8L) $\mu$ in (76 + 5.8L) $\mu$ in 0.11 <sup>o</sup>	Master Scanner and Master Plugs  Mexicali
Thread Ring Gages <sup>1,2</sup> Minor Diameter and Pitch Diameter	(0.19 to 1) in (1 to 2) in (2 to 10) in	38 $\mu$ in 40 $\mu$ in 77 $\mu$ in	LabMaster System, Gages Blocks and Sphere Probes Tijuana Juarez Monterrey
Thread Ring Gages <sup>1,2</sup> Minor Diameter Pitch Diameter Flank Angle	Up to 6.3 in Up to 6.3 in (27 to 80) <sup>o</sup>	(90 + 4.8L) $\mu$ in (90 + 4.8L) $\mu$ in 0.11 <sup>o</sup>	Master Scanner & Master Rings  Mexicali
OD Micrometers <sup>2</sup>	Up to 1 in (1 to 10) in	(5.6 + 0.6L) $\mu$ in (32 + 12L) $\mu$ in	Gage Blocks Tijuana Mexicali Juarez Monterrey Queretaro
ID Micrometers <sup>2</sup>	Up to 12 in	(32 + 12L) $\mu$ in	Gage Blocks Tijuana Mexicali Juarez Monterrey Queretaro
Calipers <sup>2</sup>	Up to 12 in (12 to 40) in (40 to 80) in (80 to 120) in	(610 + 1.1L) $\mu$ in (630 + 4.3L) $\mu$ in (940 + 0.25L) $\mu$ in (1 000 + 7.5L) $\mu$ in	Gage Blocks Tijuana Mexicali Juarez Monterrey Queretaro
Digital and Dial Indicators <sup>2</sup>	Up to 4 in	(62 + 5.6L) $\mu$ in	Gage Blocks Tijuana Mexicali Juarez Monterrey Queretaro

## Length – Dimensional Metrology

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Test Indicators <sup>2</sup>	Up to 0.06 in	(6.2 + 4.4L) $\mu$ in	Gage Blocks Tijuana Mexicali Juarez Monterrey Queretaro
Pin Gages	Up to 60 mm	3 $\mu$ m	Laser Micrometer Tijuana Mexicali Juarez Monterrey Queretaro
Laser Micrometer	Up to 60 mm	0.64 $\mu$ m	Master Plug Gages Tijuana Mexicali Juarez Monterrey Queretaro
Height Gages <sup>2</sup>	Up to 12 in (12 to 40) in	(130 + 7L) $\mu$ in (130 + 13L) $\mu$ in	Gage Blocks Tijuana Mexicali Juarez Monterrey Queretaro
Vision Equipment Optical Length <sup>2</sup>	(0 to 50) mm (50 to 100) mm	(2.1 + 0.005L) $\mu$ m (2.6 + 0.005L) $\mu$ m	Reference Glass Scale Tijuana Mexicali Juarez Monterrey Queretaro
Square/Block Flatness	Up to 4 in	9.6 $\mu$ in	Linear High Gage Tijuana Mexicali Juarez Monterrey Queretaro

## Length – Dimensional Metrology

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Distance Measurement Rulers & Flexometers <sup>2</sup>	Up to 250 mm 250 mm to 972 mm 972mm to 10 m	(3.8 + 9.2L/1 000) $\mu$ m (21 + 0.7L/600) $\mu$ m 7.8 mm	Height Gage, Distance Meter Tijuana Mexicali Juarez Monterrey Queretaro
Protractor/Angle	(Up to 90) °	2 arc min	Angle Block Set Tijuana Mexicali Juarez Monterrey Queretaro
Square/Block Parallelism	Up to 0.10 in	21 $\mu$ in	Linear High Gage Tijuana Mexicali Juarez Monterrey Queretaro
Surface Plates Local Area Flatness (Repeat Reading)	Up to 1 in	41 $\mu$ in/step	Repeat-o-Meter Tijuana Mexicali Juarez Monterrey Queretaro
Surface Plates Overall Flatness	Up to 161 in DL	(77 + 0.18 DL ) $\mu$ in	Federal Level System Tijuana Mexicali Juarez Monterrey Queretaro
Surface Finish - Source	118 $\mu$ in	2.6 $\mu$ in	Roughness Standard Tijuana Mexicali Juarez Monterrey Queretaro

### Length – Dimensional Metrology

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Surface Finish - Measure	Up to 300 $\mu\text{in}$	5.6 $\mu\text{in}$	Surface Roughness Meter Tijuana Mexicali Juarez Monterrey Queretaro
Coating Thickness Gauge <sup>3</sup>	Up to 19.84 mils (19.84 to 58.35) mils (58.35 to 202.70) mils	0.02 mils 0.09 mils 0.2 mils	Coating Thickness Standard Tijuana Mexicali Juarez Monterrey Queretaro

### Mass and Mass Related

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Pressure Source	(-14 to -0.43) psig (5.8 to 1 000) psig	0.016 % of reading + 0.000 5 psi 0.018 % of reading + 0.001 3 psi	Pneumatic Deadweight Tester Tijuana
Pressure Source	(100 to 10 000) psig	0.011 % of reading + 0.008 psi	Hydraulic Deadweight Tester Tijuana
Pressure Measuring Equipment/Measure	Up to 1 inH <sub>2</sub> O (1 to 10) inH <sub>2</sub> O (-14.7 to 300) psig (15 to 1 000) psig (1 000 to 10 000) psig (0.25 to 2.5) psig	0.006 inH <sub>2</sub> O 0.05 inH <sub>2</sub> O 0.09 psi 0.32 psi 3.8 psi 0.000 8 psi	Pressure Calibrator, Tijuana Mexicali Juarez Monterrey Queretaro
Barometric Pressure Measurement	(11.6 to 15) psia	0.011 psi	Absolute pressure module Tijuana Mexicali Juarez Monterrey Queretaro

## Mass and Mass Related

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Force - Tension/Compression	Up to 50 lbf (50 to 500) lbf Up to 1 000 lbf (1 000 to 10 000) lbf (10 000 to 50 000) lbf	0.06 lbf 1.4 lbf 2.4 lbf 26 lbf 60 lbf	Reference Load Cells Tijuana Mexicali Juarez Monterrey Queretaro
Scales/Balances <sup>2</sup>	Up to 5 g	0.021 mg + 0.6R	Class 0 Weights Tijuana Mexicali Juarez Monterrey Queretaro
	Up to 50 g (50 to 200) g 200 g to 10 kg (10 to 500) kg (500 to 1 000) kg	0.046 mg + 0.6R 0.1 mg + 0.6R 88 mg + 0.6R 4.5 g + 0.6R 110 g + 0.6R	Class F Weights Tijuana Mexicali Juarez Monterrey Queretaro
Torque – Measure Torque Tools	Up to 20 ozf·in (15 to 200) ozf·in (12.5 to 50) lbf·in (50 to 250) lbf·in (250 to 1 000) lbf·in (83.3 to 250) lbf·ft	0.5 % of reading + 0.38 ozf·in 0.25 % of reading + 0.33 ozf·in 0.33 % of reading + 0.006 lbf·in 0.31 % of reading + 0.04 lbf·in 0.31 % of reading + 0.12 lbf·in 0.31 % of reading + 1.9 lbf·ft	Torque Cell/ Torque Tester Tijuana Mexicali Juarez Monterrey Queretaro
Torque – Measure Torque Tools	(250 to 1 000) lbf·ft	0.31 % of reading + 0.88 lbf·ft	Torque Tester Tijuana Mexicali Juarez Monterrey Queretaro
Torque – Source Analyzers/Transducers	(0.07 to 0.28) Nm (0.28 to 8.47) Nm (8.47 to 16.93) Nm (16.93 to 56.44) Nm (56.44 to 67.71) Nm (67.71 to 225.7) Nm (225.7 to 1 221.05) Nm	0.054 % of reading + 0.000 18 Nm 0.053 % of reading + 0.001 5 Nm 0.015 % of reading + 0.002 1 Nm 0.007 5 % of reading + 0.022 Nm 0.002 9 % of reading + 0.021 Nm 0.007 4 % of reading + 0.025 Nm 0.002 5 % of reading + 0.02 Nm	Calibration Arms and Weights Tijuana Mexicali Juarez Monterrey Queretaro

## Mass and Mass Related

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Volumetric Calibration	Up to 5 ml (5 to 100) ml (100 ml to 250) ml 250 ml to 1.2 l (1.2 to 25) l	0.02 ml 0.074 ml 0.36 ml 0.64 ml 1.1 ml	Analytical Balance and DI Water Tijuana Mexicali Juarez Monterrey Queretaro
Air Flow	Up to 10 sccm (10 to 500) sccm (0.5 to 20) slpm (20 to 250) slpm (250 to 1 000) slpm	0.52 % of reading + 0.03 sccm 0.52 % of reading + 1.2 sccm 0.52 % of reading + 0.05 slpm 0.87 % of reading + 0.51 slpm 0.87 % of reading + 5.4 slpm	Flowmeter Calibration System Tijuana Mexicali Juarez Monterrey Queretaro
Liquid Flow Meters	(1 to 10 000 GPM DN 30 to DN 1 000	2 % of reading + 0.16 GPM	Ultrasonic Flow Meter Tijuana Mexicali Juarez Monterrey Queretaro
Air Velocity Meters	Up to 3 000 fpm	2.6 % of reading + 0.581 fpm	Anemometer Tijuana Mexicali Juarez Monterrey Queretaro

## Mass and Mass Related

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Mass	1 mg 2 mg 5 mg 10 mg 20 mg 50 mg 100 mg 200 mg 500 mg 1 g 2 g 5 g 10 g 20 g 50 g 100 g 200 g 500 g 1 kg 2 kg 5 kg 25 kg	0.13 mg 0.15 mg 0.17 mg 0.21 mg 0.24 mg 0.42 mg 0.52 mg 0.65 mg 0.74 mg 1.1 mg 1.3 mg 1.8 mg 2.4 mg 4.8 mg 12 mg 24 mg 48 mg 84 mg 0.14 g 0.25 g 0.61 g 0.6 g	Class F Weights and Balances
DOD Midas, OEM and GIDEP Sourced Procedures			Tijuana Mexicali Juarez Monterrey Queretaro
NIST 105-1 Handbook			
Up to Class F only			
Rockwell Hardness Testers	(<60) HRBW (≥60 to <88) HRBW (≥88) HRBW  (<35) HRC (≥35 to <60) HRC (≥60) HRC  (<84) HREW (≥80 to <93) HREW (≥ 93) HREW	2.1 HRBW 1.5 HRBW 1.3 HRBW  1.2 HRC 1.2 HRC 0.68 HRC  1.3 HREW 1.4 HREW 1.3 HREW	ASTM E18 Indirect Verification  Tijuana Mexicali Juarez Monterrey Queretaro

## Mass and Mass Related

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Rockwell Hardness Testers	(<65) HRKW (≥65 to <85) HRKW (≥85) HRKW  (<37) HR45N (≥37 to <66) HR45N (≥66) HR45N  (<57) HR30TW (≥57 to <70) HR30TW (≥70) HR30TW  (<78) HR15N (≥78 to <90) HR15N (≥90) HR15N  (<81) HR15TW (≥81 to <87) HR15TW (≥87) HR15TW	1.1 HRKW 1.1 HRKW 1 HRKW  1.3 HR45N 1.3 HR45N 0.75 HR45N  1.4 HR30TW 1.2 HR30TW 1.2 HR30TW  1.2 HR15N 1.1 HR15N 0.77 HR15N  1.1 HR15TW 1.1 HR15TW 1.2 HR15TW	ASTM E18 Indirect Verification  Tijuana Mexicali Juarez Monterrey Queretaro
Brinell Hardness Testers	100 HBW 10/500 142 HBW 10/3000 163 HBW 10/500 197 HBW 3000 239 HBW 10/500 248 HBW 10/3000	4.3 HBW 5.7 HBW 7.9 HBW 8.7 HBW 9 HBW 14 HBW	ASTM E10 Indirect Verification  Tijuana Mexicali Juarez Monterrey Queretaro
Micro-Indentation Hardness Testers	(100 ≤ HV ≤ 240) (240 ≤ HV ≤ 600) HV > 600	24 HV 28 HV 35 HV	ASTM E92 Indirect Verification  Tijuana Mexicali Juarez Monterrey Queretaro

## Mass and Mass Related

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Durometers Force Types A, B, O Types C, D, DO Types M, OO	(56.08 to 820.87) gf Up to 4530 gf Up to 113 gf	0.99 grf 6.2 grf 0.58 grf	ASTM D2240 – Section 7 Tijuana Mexicali Juarez Monterrey Queretaro
Indenter Length Indenter Angle Indenter Radius	Up to 20 mm Up to 35° Up to 1 mm	2.4 um 0.2° 3.3 um	

## Photometry and Radiometry

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Optical Power Source (850, 1 300, 1 550) nm	(-7 to 0) dBm	0.35 dB	Laser Source Tijuana Mexicali Juarez Monterrey Queretaro
Optical Power Measure-Linearity Measure –Accuracy 800 nm to 1650 nm	(-80 to 10) dBm Up to 10 mW	0.035 dB 3.5% of reading + 0.073 uW	Optical Power Sensor Tijuana Mexicali Juarez Monterrey Queretaro
Optical Attenuation Source (1 300, 1 550) nm	(-60 to 0) dB	0.13 dB	Optical Attenuator Tijuana Mexicali Juarez Monterrey Queretaro

## Thermodynamic

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Humidity – Source <sup>4</sup>	43 %RH 75 %RH 97 %RH	1.3 %RH	Indicator/Hygrometer Saturated Salt Baths Tijuana Mexicali Juarez Monterrey Queretaro
Humidity - Measure	(>0 to 99) %RH	1.3 %RH	Indicator/Hygrometer Tijuana Mexicali Juarez Monterrey Queretaro
Radiation (Infrared) Thermometers	(50 to 100) °C (100 to 300) °C (300 to 500) °C	1.7 °C 5.3 °C 8.2 °C	Blackbody Source (flat plate) $\epsilon = 0.95, \lambda = (8 \text{ to } 14) \mu\text{m}$ Tijuana Mexicali Juarez Monterrey Queretaro
Radiation (Infrared) Thermometers	(100 to 1 200) °C	21 °C	High Temperature Blackbody Source (cavity) $\epsilon = 0.995, \lambda = (8 \text{ to } 14) \mu\text{m}$ Tijuana Mexicali Juarez Monterrey Queretaro
Temperature Source- Measuring Equipment	(-30 to 0) °C (0 to 150) °C (150 to 300) °C (350 to 1 200) °C	0.26 °C 0.23 °C 0.53 °C 5.7 °C	Reference Thermometer w/ PRT Tijuana Mexicali Juarez Monterrey Queretaro

## Thermodynamic

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Temperature Source / Measure	(-200 to 0) °C (0 to 100) °C (100 to 300) °C (300 to 420) °C (420 to 650) °C	0.03 °C 0.05 °C 0.07 °C 0.11 °C 0.14 °C	Reference Thermometer w/ PRT Tijuana Mexicali Juarez Monterrey Queretaro
Temperature Source / Measure	0 to 1 200 °C	0.6 °C	Type S Reference Thermocouple Mexicali Juarez Monterrey Queretaro
Temperature Measurement	(650 to 1 200) °C	0.1 % of reading + 1.4 °C	Type R Reference Thermocouple Mexicali Juarez Monterrey Queretaro
Dew Point	(-40 to 60) °C	2.5 °C	Reference Dew Point Indicator Tijuana Mexicali Juarez Monterrey Queretaro

## Time and Frequency

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Chronometers, Stopwatches, Timers	1 ms to 100 000 s	0.12 ms	Frequency Counter Tijuana Mexicali Juarez Monterrey Queretaro

## Time and Frequency

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Time - Source	1 ms to 100 000 s	(4.8 x 10 <sup>-3</sup> ) µs	Frequency Counter Tijuana Mexicali Juarez Monterrey Queretaro
Frequency – Measure <sup>2</sup>	150 kHz to 1.3 GHz	(2.4 x 10 <sup>-8</sup> ) Hz + 2R	Measuring Receiver Tijuana Mexicali Juarez Monterrey Queretaro
Frequency – Measure <sup>2</sup>	DC to 3.2 GHz (3.2 to 20) GHz	(9.4 x 10 <sup>-6</sup> ) Hz + 2R 2.6 Hz	Frequency Counter Tijuana Mexicali Juarez Monterrey Queretaro
Frequency - Source	(0.01 to 120) Hz 120 Hz to 1.2 kHz (1.2 to 120) kHz 120 kHz to 1.2 MHz (1.2 to 2) MHz	3 µHz/Hz + 61 µHz 3 µHz/Hz + 61 µHz 3 µHz/Hz + 120 µHz 3 µHz/Hz + 10 mHz 3 µHz/Hz + 12 mHz	Multiproduct Calibrator Tijuana Mexicali Juarez Monterrey Queretaro
Frequency – Source <sup>2</sup>	2 MHz to 6 GHz	1.3 µHz/Hz + 0.02 Hz	Signal Generator Tijuana Mexicali Juarez Monterrey Queretaro
Frequency – Source <sup>2</sup>	10 MHz to 26.5 GHz	(1.2 x 10 <sup>-7</sup> ) Hz + R	Signal Generator Frequency Counter Tijuana Mexicali Juarez Monterrey Queretaro

### Time and Frequency

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Tachometer (Stroboscope)	Up to 100 rpm (100 to 1 000) rpm (1 000 to 99 999) rpm	0.03 % of reading + 0.014 rpm 0.03 % of reading + 0.14 rpm 0.03 % of reading + 1.4 rpm	Tachometer/Stroboscope Tijuana Mexicali Juarez Monterrey Queretaro
Discharge Time	Up to 999.9 s	2 % of reading + 0.14 sec	Charge Plate Analyzer Tijuana Mexicali Juarez Monterrey Queretaro

